

Certificate of Analysis SALSA® MLPA® Probemix P114 Long-QT

Catalogue #	P114-025R, P114-050R, P114-100R	
Product name	Probemix P114 Long-QT	
LOT	C1-0519	
Σ	25, 50, or 100 reactions.	
Shipping conditions	Dry ice or cooling elements.	
X	Store upon arrival between -25°C and -15°C.	
	Expiration date: May 2024, when stored at recommended conditions. This not be frozen/thawed more than 25 times.	s product should
Purpose	This product has been developed to determine the DNA copy numbers of human KCNQ1, KCNH2, KCNJ2, KCNE1, and KCNE2 genes, as described of the product description. This probemix is designed for use only in c SALSA MLPA reagent kits and Coffalyser.Net analysis software as descri General Protocol.	in table 1 and 2 ombination with
Quality control specifications	- Sufficient distance between peaks, absence of extra or shoulder peaks, and completeness of hybridisation of each individual probe, as tested on Applied Biosystems and Beckman/SCIEX GeXP sequencers.	Test result
	 Standard deviation of each individual probe ≤0.10, when tested on 23 different DNA samples of healthy individuals, extracted by various methods. Each individual probe meets reaction-specific criteria when tested on a single DNA sample under various experimental conditions. 	
	 No-DNA controls result in only five major peaks shorter than 121 nucleotides (nt): four Q-fragments at 64, 70, 76 and 82 nt, and one peak in the range of 0-40 nt corresponding to the unused portion of the fluorescent PCR primer. Non-specific peaks longer than 121 nt AND with a height <25% of the median of the four Q-fragments are not expected to affect MLPA reactions when sufficient (50-250 ng) sample DNA is used. Note: We observed one prominent peak below the 25% threshold with a length of approximately 142 nt in a No-DNA control. 	PASS

None of the ingredients are derived from humans, animals, or pathogenic bacteria. Based on the concentrations present, none of the ingredients are hazardous as defined by the Hazard Communication Standard. A Safety Data Sheet (SDS) is not required for these products: none of the preparations contain dangerous substances (as per Regulation (EC) No 1272/2008 [EU-GHS/CLP] and amendments) at concentrations requiring distribution of an SDS (as per Regulation (EC) No 1272/2008 [EU-GHS/CLP] and 1907/2006 [REACH] and amendments). If spills occur, clean with water and follow appropriate site procedures.

More information: www.mrcholland.com; www.mrcholland.eu	
•••	MRC Holland bv; Willem Schoutenstraat 1 1057 DL, Amsterdam, The Netherlands
E-mail	info@mrcholland.com (information & technical questions) order@mrcholland.com (orders)
Phone	+31 888 657 200



Certificate of Analysis

SALSA MLPA Probemix P114-C1 Long-QT sample picture




Figure 1. Capillary electrophoresis pattern from a sample of approximately 50 ng human male control DNA analysed with SALSA MLPA Probemix P114 Long-QT (C1-0519).

This lot was certified by MRC Holland on 25 September 2019.

This certificate is a declaration of analysis at the time of the manufacturing process. All assays were run in compliance with manufacturer's instructions for use.

Implemented changes in the COA

Version 02 - 03 March 2021 (6)

- COA restructured and adapted to a new template.

Version 01 - 25 September 2019 (04)

- Not applicable, new document.